

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.usplo.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/652,636	08/31/2003	Lendell Martin SR.	MART 006	1829
Com Ma Chan a	7590 05/04/200	7	EXAMINER	
Guy McClung # 114		JASTRZAB, KRISANNE MARIE		
5315-B F.M. 1 Houston, TX 7			. ART UNIT	PAPER NUMBER
110451011, 171 /	7005 1110		1744	
				DEL MEDICAGO
			MAIL DATE	DELIVERY MODE
			05/04/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	Application No.	Applicant(s)	
	10/652,636	MARTIN, LENDELL	
Office Action Summary	Examiner	Art Unit	
·	Krisanne Jastrzab	1744	
The MAILING DATE of this communication ap Period for Reply	pears on the cover sheet w	ith the correspondence addres	:s
A SHORTENED STATUTORY PERIOD FOR REPL WHICHEVER IS LONGER, FROM THE MAILING D. - Extensions of time may be available under the provisions of 37 CFR 1. after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period Failure to reply within the set or extended period for reply will, by statut Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	DATE OF THIS COMMUNI 136(a). In no event, however, may a will apply and will expire SIX (6) MOI e. cause the application to become A	CATION. reply be timely filed NTHS from the mailing date of this community BANDONED (35 U.S.C. § 133)	
Status			
 1) ⊠ Responsive to communication(s) filed on 21 F 2a) ⊠ This action is FINAL. 2b) ☐ This 3) ☐ Since this application is in condition for allowed closed in accordance with the practice under the condition of t	s action is non-final. ance except for formal mat		rits is
Disposition of Claims			
4) ☑ Claim(s) 19-32 is/are pending in the application 4a) Of the above claim(s) is/are withdra 5) ☐ Claim(s) is/are allowed. 6) ☑ Claim(s) 19-32 is/are rejected. 7) ☐ Claim(s) is/are objected to. 8) ☐ Claim(s) are subject to restriction and/or	wn from consideration.		
Application Papers			•
9) The specification is objected to by the Examine 10) The drawing(s) filed on is/are: a) accomposed and applicant may not request that any objection to the Replacement drawing sheet(s) including the correct to by the Examine and the specific properties of t	cepted or b) objected to drawing(s) be held in abeyant ition is required if the drawing	nce. See 37 CFR 1.85(a). (s) is objected to. See 37 CFR 1.	
Priority under 35 U.S.C. § 119			٠
12) Acknowledgment is made of a claim for foreign a) All b) Some * c) None of: 1. Certified copies of the priority document 2. Certified copies of the priority document 3. Copies of the certified copies of the priority document application from the International Burea * See the attached detailed Office action for a list	ts have been received. ts have been received in A prity documents have been u (PCT Rule 17.2(a)).	opplication No received in this National Stag	je
Attachment(s) 1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date	Paper No(Summary (PTO-413) s)/Mail Date nformal Patent Application 	

DETAILED ACTION

It is noted that two claims numbered "31" were submitted in the amendment filed 2/21/2007. The second recited claim has been renumbered "32" according to Rule 126.

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 25, 26 and 32 (second claim recited as 31 in amendment) are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

With respect to claim 25, this claim is found to be vague and indefinite because it sets forth "an interior surface", while claim 19 from which it depends previously set for "an interior surface". As such it is unclear as to whether claim 25 is attempting to further limit the previously recited surface or setting forth another surface. Clarification is required. Also, "the UV resistant material" lacks proper antecedent basis.

With respect to claim 26, "the UV resistant material" lacks proper antecedent basis.

With respect to claim 32 (the second recited claim 31), this claim is found to be vague and indefinite because it refers to "the method of claim 19", however claim 19 is an apparatus claim. Clarification is required.

Claim Rejections - 35 USC § 102

Art Unit: 1744

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 19-21 and 28-32 are rejected under 35 U.S.C. 102(b) as being clearly anticipated by Hines, Jr. et al., U.S. patent No. 4,129,013.

Hines, Jr. et al., teach an enclosure for an HVAC coil element and wherein a dual pan configuration is employed with the bottom pan including rib means and feet to space the pan from the base of the primary pan. The enclosure is formed of a plurality of interlocking panels. See particularly column 5, lines 40-55 and Fig. 11.

Claim Rejections - 35 USC § 103

Claims 22-27 are rejected under 35 U.S.C. 103(a) as being unpatentable over Hines, Jr. et al., as applied to claims 19-21 and 28-32 above, and further in view of either of Elmore U.S. patent No.'s 6,234,241 B1 or 5,558,158.

Both Elmore patents teach hygienic air plenum system formed from a containment means having an inner and outer surface with the inner surface coated with a protective, UV resistant coating. The coating can be non-porous materials chosen from metal, foil, a laminate such as formica, acrylic, plastic, glass, etc. The containment means can used for the coil of an air conditioning unit, and also contains a UV source. The construction provides optimum stability with minimal prospective bacterial growth. See column 2, lines 35-65, column 4, lines 35-65, and column 5, lines 23-48 of either.

Art Unit: 1744

It would have been obvious to one of ordinary skill in the art to construct the enclosure of Hines, Jr. et al., from the materials taught in the Elmore patents and to include the UV light source as well, because it would optimize the structural stability while preventing bacterial growth within the system.

Response to Arguments

Applicant's arguments with respect to claims 19-32 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Krisanne Jastrzab whose telephone number is 571-272-1279. The examiner can normally be reached on Mon.-Thurs. 6:00am-4:30pm.

Application/Control Number: 10/652,636

Art Unit: 1744

Page 5

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Gladys Corcoran can be reached on 571-272-1214. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Krisanne Jastrzab Primary Examiner Art Unit 1744

May 2, 2007